



PROFESSIONAL USER'S MANUAL

COAG-SENSE®
PROTHROMBIN TIME (PT)/INR
MONITORING SYSTEM



No part of this publication may be reproduced, transmitted, transcribed, stored in a retrieval system or translated into any language or computer language, in any form or by any means, including, but not limited to, electronic, magnetic, optical, chemical, manual, or otherwise without written permission of CoaguSense. CoaguSense has made every reasonable effort to ensure that all the information contained in this manual is correct at the time of printing. However, CoaguSense reserves the right to make any changes necessary without notice as part of ongoing product development.

If you have any questions or concerns with the Coag-Sense® Prothrombin Time (PT)/INR Monitoring System, please contact CoaguSense Technical Support at:

CoaguSense, Inc.

48377 Fremont Blvd., STE. 113

Fremont, CA 94538

USA

Toll Free: 1-866-903-0890

E-Mail: techsupport@coagusense.com

© Copyright 2022, CoaguSense, Inc. All rights reserved.

Coag-Sense is a registered trademark of CoaguSense, Inc.

Table of Contents

Table of Contents	1
1. Introduction	1
2. System Description	3
3. Meter Specifications	10
4. Performance Characteristics	12
5. Warnings and Precautions	13
6. Hazards and Symbols	17
Directions for Use	19
7. Meter Setup	19
8. Performing a Control Test	29
9. Collecting a Fingertick Sample	34
10. Performing a PT Test	37
11. Performing a STAT test	43
12. Reviewing the Memory	47
13. Printing	50
14. Network Connectivity and Security	52
15. Meter Software Update	53
16. Battery	54
17. Cleaning and Disinfecting the Meter	55
18. Troubleshooting	57
19. Warranty	65
20. Reordering Information	67
21. EMC Tables	68
22. Index	71

1. Introduction

The Coag-Sense® Prothrombin Time (PT)/INR Monitoring System

Intended Use

The Coag-Sense Prothrombin Time (PT)/INR Monitoring System is an *in vitro* diagnostic device that provides quantitative prothrombin time (PT) results, expressed in seconds and International Normalized Ratio (INR) units. It uses fresh capillary whole blood. It is intended for use by health care professionals at the point of care to monitor patients who are on warfarin-type (coumarin) anticoagulation therapy. The device is not intended to be used for screening purposes.

Importance of PT/INR Monitoring

Blood-Clotting Time:

The rate at which blood clots is measured in units is called International Normalized Ratio (INR). It is very important for patients to stay within their individual target INR range. If the INR is too low, the risk of blood clots increases. If the INR is too high, the risk of hemorrhaging increases. The patient's physician will determine the most appropriate INR range for the patient, depending upon the patient's indication and how they respond to the oral anticoagulants.

Anticoagulation Medication:

Oral anticoagulation medications, are typically prescribed to patients to avoid unwanted clots. The patient's blood clotting time must be monitored to ensure that their dosage is correct.

Oral anti-coagulation medication is prescribed to patients with acute and chronic conditions including, but not limited to: congestive heart failure, atrial fibrillation, prosthetic heart valve, myocardial infarction, joint replacement, deep vein thrombosis, pulmonary embolism, thrombotic stroke, coronary artery disease, venous thromboembolism and cancer.

Important Information Regarding Manual

The purpose of the Coag-Sense Prothrombin Time (PT)/INR Monitoring System User Manual is to help you understand your Coag-Sense PT/INR system, its parts, and its intended function. It provides you with the information you need to perform a PT test with the Coag-Sense PT/INR system.

You must complete proper training on the Coag-Sense PT/INR system and configure the meter to your needs before you begin using the system. It is also important to read this entire User Manual and the inserts that come with the disposable Coag-Sense test strips. This User Manual has different formats and symbols to distinguish warnings, notes, and meter buttons.



WARNING: This indicates a warning or precaution.

Please read and understand all warnings and precautions. They tell you about potential safety hazards and situations that may cause injury. If you have any questions, please contact CoaguSense Technical Support at 1-866-903-0890 (USA).

2. System Description

The Coag-Sense Prothrombin Time (PT)/INR System is used for quantitative measurement of INR (International Normalized Ratio) based on a Prothrombin Time (PT) response to monitor the effect of therapy with vitamin K antagonists like Coumadin® (warfarin). The system uses fresh, capillary whole blood.



Meter:

The meter has a TFT color LCD **Touch screen** that shows results, information, icons and results recalled from memory. To select an option, gently click on display button. There are three touch buttons, **Cancel or Previous Screen Button, Home Screen Button and View Menu Button Screen** for the Guided User Interface (GUI) operation. The **Power ON/OFF** button is located on the right side of the meter. The **NFC (Near Field Communication) Tag scanner** is a built-in scanner that is used to scan the NFC card containing the strip (Control and Test Strip) data. **Strip Insertion Area** guides the test strip into the meter. **Micro USB/Power Adapter port** is a micro-USB port used to plug to the power adapter. Multipurpose **USB port** can be used to connect the meter to a) portable printer, b) computer, c) barcode scanner or other Coag-Sense approved accessories. **Ethernet Port** is used to connect the Ethernet cable for a wired connection, this port is provided with a port cover. **Reset button (enclosed within the ethernet port cover)** is used to reset the meter in case of software or power-cycle issues.

The meter performs a self-check when it is first powered ON and every time a test strip is inserted. If there are any problems detected during self-check, an error message is displayed on the touchscreen. Refer to the “Troubleshooting” section of this manual or contact Technical Support for assistance.

Test Strips:

A test strip is inserted and heated in the meter prior to sample application. The strip contains a tiny wheel with spokes that pull the sample into the reaction well. The spokes quickly and completely mix the sample with the clot initiating component of the test strip.

The PT time is determined from when the sample is drawn into the reaction well of the test strip and is detected by a beam of light until a clot forms and interrupts another beam of light. The PT seconds result (a true prothrombin time) is converted to an INR (International Normalized Ratio) using the INR normalization data communicated by the NFC tag and subsequently stored in the meter. INR is a mathematical correction of the

PT result that adjusts for sensitivity differences among different PT/INR systems.



Control Strips and Control Activation Solution

Quality control is an important part of PT testing to verify the integrity of the performance characteristics of the testing system. The Coag-Sense Meter has been designed with multiple internal systems to ensure proper system function. When powered ON, the meter runs an extensive self-check protocol to ensure, for example, that operating temperature, timing functions, battery level and optical and mechanical functions are within specification. There are 2 Low and High Control Strips, and a Control Activation Solution shipped with each test strip kit. Each control strip contains plasma which is generated from a pool of normal donors where the Vit K dependent proteins are removed and added back at different levels to represent the 'Low' and 'High' level ranges. Real plasma allows for a fully functional liquid quality control test of both the a) reagent's ability to generate a clot and b) the analyzer's ability to detect a clot. Control testing confirms the performance of the system and should be completed immediately for each new lot of test strips received

NFC Tag:

Near Field Communication (NFC) Tag is a micro data tag with antenna that contains the lot specific test strip kit information. Transfer of the lot specific data to the meter can be accomplished through surface contact of the matching NFC symbols. The NFC tag communicates the unique data for each lot of strips to the meter. The meter reads the data stored in

the NFC Tag and auto populates the relevant test strip information on the screen.

In the absence of NFC Tag, the user may manually enter the lot and Barcode number present on the strip packaging using the keypad on the touch screen. A stylus with a rubber capacitive tip may be used to facilitate typing.

Power Supply and Battery:

Coag-Sense PT/INR Professional System can be operated only with the power adapter provided. The power adapter also serves as a charger. It charges the built-in Lithium Polymer Battery. The battery life is shown on top right corner of the meter.

Note: *The battery is not user replaceable.*

To save power, if left unattended for a set time (user configurable), the meter will enter sleep mode. To power meter OFF, manual press and hold of the Power Button is required. The meter retains all results obtained up to that point in its memory.

Coag-Sense PT/INR Professional System (**Catalog # 03P70-02**) is supplied with the following items:

Catalog Name	Qty
Coag-Sense PT/INR Meter	1
Coag-Sense PT/INR System Professional User's Manual	1
Coag-Sense PT/INR System Professional Quick Reference Guide	1
A/C Micro USB Power Supply	1
Sample Transfer Tubes with Plungers	1 pk
Single-Use, 21g Auto Safety Lancets (sample pack)	1 pk
Carrying Case	1

To perform a test, you require the following:

- Coag-Sense Test Strip Kit- 50 (**Catalog # 03P56-50**) is supplied with the following items:


Item Description	Qty
Patient Test Strips	50
Low Control Strips	2
High Control Strips	2
Control Strip Activation Solution	1
Lot Info label w/NFC Tag	1
Sample Transfer Tubes w/ plungers	1 pk
Package Insert	1

- Following are standard medical supplies that are required with each use:
 - Gauze
 - Isopropyl alcohol or alcohol wipes,
 - Single-Use- 21g Auto Safety Lancets,
 - Puncture-resistant bio-hazard (SHARPS) container

Note: These materials are not provided with the PT/INR system. The Coag-Sense Professional Test Strip Kit- 50 may be ordered from your distributor separately.


Overview of Buttons and Icons

The buttons and icons that appear during normal operation are shown here, along with their respective meanings. Error messages and their descriptions are provided in 'Troubleshooting' section.

Buttons/Icons	Meaning
	Power ON/OFF To power ON the meter, press and hold Power Button. To Enter/Exit Sleep Mode press the button once quickly and press the button again and hold for few seconds.
	Cancel or return to previous screen
	Go to the home screen
	View additional menu
	Common Keypad input: a and b) is the input completion button. Returns to previous screen when selected. c) Change language button. Enables the user to select keyboard language.
	The Status bar: a) Date and Time b) Sound ON/OFF status c) Alarm Status d) Bluetooth ON/OFF status e) Wi-Fi ON/OFF status f) Battery level

Buttons/Icons

Meaning



Icons on the touchscreen:


a) Back Icon- Directs to the previous screen

b) Home Icon- Directs to the home screen

c) Settings Icon- Directs to the setting screen

d) User Information Icon- Directs to the User Information screen

e) Log-out Icon- Directs to the logout pop up screen.



Home Screen:

a) Test Icon- Directs User to Test Strip information screen and subsequently to testing screen

b) Control Icon- Allows User to select between High and low control tests. Then directs User to Test Strip information screen and subsequently to control screen.

c) Results- Allows User to view Patient and Control test results. Refer to Reviewing the memory section for detailed information

d) Settings- Directs User to the Settings menu.

3. Meter Specifications

Operating Temperature	65°F to 90°F (18°C to 32°C)
Operating Humidity	10% to 90% (without condensation)
Storage Temperature	32°F to 122°F (0°C to 50°C)
Storage Humidity	20% to 80% (without condensation)
Altitude	10,000 ft (3,048 m) above sea level
Memory	Capable of storing up to, • 2000 patient test results with date and time • 500 control test results with date and time • 1000 Operator accounts
Lithium Battery	Rechargeable Lithium Polymer Battery (3.7V, 2350mAh)
Battery Capacity	Fully charged (6 hours of charging) can run ~200 tests
Power Input	120V AC Adapter (Use with Coag-Sense Adapter Only)
AC Input	100-240V~, 50-60Hz, 0.5A (Mains supply voltage fluctuation: $\pm 10\%$)
Power Output	5.0V, 2.0A
Pollution Degree	2
Overvoltage Category	II
Use Circumstance	Indoor only
Blood Sample Size	10-12 μL
Communication Port	Micro and Standard USB
Size in mm (Height x Width x Depth)	152 x 100 x 29.5
Weight in g	315g

Equipment Classification	Class II with external power supply. Internally powered when operated with battery. IPX0 rating.
 WARNING: Use the Coag-Sense meter along with the provided Power Adapter only. Use only AC adapter type UES12ICP-050200SPA (Manufacturer: Dongguan Shilong Fuhua Electronics).	

4. Performance Characteristics

Expected Values: Results are reported in INR units equivalent to the plasma reference method. For PT testing, variations in the source of thromboplastin may cause some differences in results between methods. It is recommended that the same method be used to monitor the anticoagulation therapy over time.

Measuring Range: INR 0.8 to 8.0

Normal Range: The following example represents a common normal range for the Coag-Sense PT/INR system.

INR: 0.8 to 1.2

PT: 11.6 to 14.5

5. Warnings and Precautions



This test system is not recommended for patients who have recently taken or are currently taking any type of Heparin anticoagulants. The system should also not be used to monitor patients on direct oral anticoagulants (DOACs) including Factor Xa and Direct Thrombin inhibitors.

Test Site and Blood Sample

- The Coag-Sense PT/INR system is for in vitro diagnostic use only.
- The Coag-Sense meter will not produce a result if the test strip is past its expiration date.
- The quality of the blood sample can affect PT test results. A blood sample of poor quality can produce unreliable results. Read the section on **“Collecting a Fingerstick Sample”** for more information.
- Blood samples must be applied to the test strip **immediately** after collection or the blood begins to clot, causing unreliable results.
- The blood sample transferred to the test strip must be a minimum of 10 µL in volume. Low sample volume may cause an error message.
- Use only fresh fingerstick capillary blood for testing. The blood should only come in contact with the Sample Transfer Tubes provided with the Coag-Sense PT/INR system. Other products may have anti-coagulant agents on their surfaces and result in unreliable test results.
- Squeezing the fingerstick site excessively (milking) releases interstitial “tissue layer” fluid that can cause unreliable results.
- The fingerstick site can be washed with warm water and soap, and then completely dried. The site must be clean of all hand oils/lotions and foreign matter, which may cause unreliable results.
- **If Isopropyl Alcohol (IPA) wipes are used, wipe the fingerstick site with a gauze pad and make sure the site is completely dry.**

If any alcohol remains (or is re-introduced) on the finger, it may cause unreliable results.

- Do not use wipes containing chlorhexidine gluconate, as it may produce unreliable results.
- **The quality of fingerstick and the sample delivery technique are important to the test results. If there is a question about the sample or sample collection, obtain a new strip, repeat the fingerstick on a different finger, and test again.**
- If you need to repeat a test, use a different finger for the fingerstick, since blood may have started to clot on the first finger, which may cause unreliable results.
- If there is a bubble or an air pocket showing in the blood sample in the collection tube, start the test over. Use a new fingerstick (using a different finger and collection tube) or results may be unreliable.

Meter

- The meter has a built-in rechargeable lithium polymer battery (3.7V, 2350 mAh).
- Use only the included Coag-Sense Power adapter with the meter or damage to the meter may result.
- The meter shall be in the position that it is easy to disconnect power.
- The meter is a delicate instrument, and should be handled with care. Dropping or other mishandling may cause damage to the meter. If this should occur, call Technical Support.
- Do not allow any liquids to spill on the meter. If this should occur, call Technical Support.
- Do not put the meter in liquid. Do not allow liquids to get into any of the connectors or plugs on the meter.
- Only use the method provided in this User Manual to clean the Coag-Sense PT/INR meter.

- Do not move or touch the meter while it is running a test. Unreliable results may occur.
- Do not pull the strip out during a test while the wheel is spinning. STOP the test, meter will display "Test Cancelled, Remove Strip" The strip should be removed at this time only.
- Store and use the Coag-Sense PT/INR system following the instructions in this manual.
- This equipment is tested to meet the limits for medical devices, which are designed to provide a reasonable protection against harmful interference when the equipment is operated in a clinical or home environment. If not installed and used in accordance with these instructions, it may cause harmful interference to other devices in the vicinity. If this equipment does cause harmful interference to other devices, which can be determined by turning the equipment on and off, the user is encouraged to try to correct the interference by one or more of the following measures:
 - Reorient or relocate the receiving device.
 - Increase the separation between the equipment.
 - Connect the equipment to an outlet on a circuit different from that to which the other devices are connected.
- Any equipment connected to the data port must be certified to IEC 60601-1. If you connect any equipment that is not recommended by CoaguSense, you are responsible for meeting the requirements of this standard.
- In the unlikely event of an electric power surge (i.e., severe static discharge during a thunderstorm), when using the power adapter, the display screen may go blank. If this occurs, unplug the power supply from the back of your meter, wait 5 seconds and plug it back in. Normal operation should return, but you may have to reset the time and date.
- DO NOT OPEN THE METER. Do not attempt to repair or modify this meter. The Coag-Sense meter does not require any periodic maintenance and there are no user serviceable parts inside. If you have problems, please contact Technical Support. The Coag-Sense

Prothrombin Time (PT)/INR Monitoring System needs special precautions regarding EMC and needs to be put into service according to the EMC information provided in this manual.”

- “Portable and mobile RF communications equipment can affect The Coag-Sense Prothrombin Time (PT)/INR Monitoring System.”
- “The use of accessories, transducers and cables other than those specified by CoaguSense, may result in increased EMISSIONS or decreased IMMUNITY of the EQUIPMENT.”
- “The Coag-Sense Prothrombin Time (PT)/INR Monitoring System should not be used adjacent to or stacked with other equipment and that if adjacent or stacked use is necessary, the Coag-Sense Prothrombin Time (PT)/INR Monitoring System should be observed to verify normal operation in the configuration in which it will be used.”

Test Strips/Control Strips/Control Strip Activating Solution

- The test strips are designed for single use only. Do not reuse the test strips.
- Patient samples, controls, used strips, transfer tubes and lancets are potentially infectious. Discard used materials in a puncture resistant, biohazard waste container using universal precautions.
- PT Test Strips, Control Strips, and Control Activation Solution are perishable goods with a limited shelf life. Do not use any of these items if the expiration date has passed.
- Refer to the package insert that is supplied with each box of test strips for more information.

6. Hazards and Symbols

	Warning: This indicates a warning or precaution, requiring special attention.
	Class II Equipment. The Power Adapter is double insulated.
	Biological Risks: Disposable items pose biological risks. The strips and fingerstick materials should be disposed of in appropriate biohazard waste containers.
	Electronic device. Dispose of unit properly.
	Use by/Expiration Date
	Lot number
	For In vitro diagnostic use
	Storage temperature range
	Manufacturer
	Single Use Only – Do Not Reuse
	This product meets the provisions of Council Directive 98/79/EC for In Vitro Diagnostic Devices.
	Authorized/European Representative
	Catalog Number
	The system fulfills the U.S safety requirements (NEMKO listed)
	Consult Instructions for Use



Near Field Communication (NFC) Tag



Serial Number



Contains a Unique Identification
Number



Bluetooth Connectivity, BLE Enabled
Devices



Wireless (WIFI) Network



To identify the country of manufacture
of products

Directions for Use

7. Meter Setup

Operating Conditions

To ensure that your Coag-Sense PT/INR system is working correctly, be sure the following conditions are met:

- Be sure that the meter and strips are at room temperature before use. Operating conditions are between 65°F and 90°F (18°C and 32°C). **The meter will not allow a test to proceed unless it is within the operating temperature range.**
- Relative humidity should be between 10% and 90%, without condensation, for testing.
- Avoid dropping the meter or treating it roughly.
- Use the meter only on a level, stable surface.
- Do not move the meter during testing.

Power ON/OFF

- The Coag-Sense PT/INR Professional System can be operated with the power adapter provided. The power adapter also serves as a charger.
- Place the meter on a flat, stable surface. To turn the meter ON/OFF, **press and hold** the POWER button on the right side of the meter at least 5 to 7 seconds until screen goes dark.


System and User Settings:

The meter is set to default factory settings, English is the default language and time/time zone is Pacific Standard time (UTC-8:00). User may modify


User settings as appropriate. Refer to User Settings section in this manual for the list of settings and their functions. These User settings help the User to configure their PT/INR meter. **Note that all Settings options will only be available when logged in as Admin.**





When logged in as a standard user only




Device Information and Device Setting will be displayed.








No.	Action	Image guided instruction
Device Information Setting		
1.	Device Information Screen	A screenshot of the 'Device information' screen in the application. The status bar at the top shows '01/01/2018 12:01 AM' and various icons. The screen title is 'Device information'. It lists the following details: 'Device name' (Coag-Sense PT/INR System), 'Serial number' (D001339G0362), 'Manufacturer' (CoaguSense, Inc.), 'Software version' (1.1.8), 'Test module version' (1.0.9), 'Build version' (1.0.4), 'Kernel version' (4.1.15+ga0b5a61), 'Hardware version' (1.0.9), 'MAC address (Wired)' (F8:DC:7A:16:3D:96), and 'MAC Address (Wireless)' (00:25:CA:14:42:FC). Each version field has a right-pointing arrow.




No.	Action	Image guided instruction
2.	Software Version Update	 10/15/2018 4:33 PM Version info. New version 1.0.7.3 2018/03/04 17:36 Version info. 95 % Cancel 2018/03/04 17:41 Update Rebooting device is required for the update. Do you want to reboot now? Reboot Not now
Device Settings		
1.	Device Settings Screen	 05/09/2018 3:22 pm Device settings Theme Color Language English Date Format 01/15/2018 Time Format 3:30 pm Date/Time Time zone Brightness Sound Target range Low 1.1 / High 2.2 Alarm






No.	Action	Image guided instruction
2.	Theme: User chooses from Mono/Blue/Color	
3.	Language Setup <i>Note: The default language is set to English.</i>	
4.	Date Format Setup	
5.	Time Format Setup	

No.	Action	Image guided instruction
6.	Date and Time Setup <i>Note: The DST option must be turned on to automatically adjust for daylight savings time.</i>	
7.	Time Zone Setup	
8.	Brightness: User chooses from range of 1-7	
9.	Sound Setting: User chooses to Turn sound ON or OFF	

No.	Action	Image guided instruction
10.	Target Range Setup	
11.	Alarm: User can set alarm as a reminder to test at a certain day and time of the week. User may edit or delete an alarm at any time.	
Communication Settings		
1.	The screen lists the status of the communication channels. If connected to Wireless (Wi-Fi), the touchscreen displays the name of the Wi-Fi network. The forward button on each type will direct the User to a detailed view.	
2.	Wireless (Wi-Fi) Setting: This screen displays icons that a) Scans for Wireless networks nearby b) Wireless ON/OFF icon Clicking the 'Connect' icon will require User to input password to connect to the network.	

No.	Action	Image guided instruction
3.	Upon clicking the 'Setup' icon, the touchscreen displays the Wireless Information screen, User may a) select a connection method and b) input the fields containing network information. Click 'Confirm' to proceed and input the password for connecting to a network. If the desired wireless network cannot be automatically connected via DHCP then it can be configured manually.	
4.	Wired Setting: Setup wired network connection by inputting the fields displayed in the screen. <i>Note: If the connection method is DHCP, the contents are automatically assigned if the network with the DHCP server is plugged in.</i>	
5.	Broadcast Bluetooth will pair the Coag-Sense PT/INR meter to an external device. <i>Note: Set the external device ready to connect to the Coag-Sense meter.</i>	


No.	Action	Image guided instruction
6.	The touchscreen displays a message if the Bluetooth connection fails.	
Note: The screen shot on the right shows the Bluetooth Connectivity icons.		
External Connection		
7.	a) Screen lists server IP address for socket communication b) Open/Close port for serial communication with external program	
Admin Settings		
8.	QC Lock Settings: If a Control test (High Control or Low Control) does not produce a result, or if the result is outside of the target value range, the meter can be locked from further patient testing. The frequency for required testing can be set.	
Note: Refer to QC Lockout section below for detailed information.		

No.	Action	Image guided instruction
9.	Sleep Settings: If left unattended, for a set of time (User configurable), the meter will enter to sleep mode. To power meter OFF, manual press and hold of the Power Button is required.	
10.	User Management: User Management screen, a) displays the 'Admin' account that was previously set up (Initial Setting Section). Admin User may choose to change the password by clicking this field. b) displays the list of User accounts. c, d) lets you add and delete user accounts.	 Edit Administrator Login 
11.	Create and Edit User Accounts: You can Create User accounts or Edit User accounts and enter their department information.	 

QC Lockout (Optional Settings):

When the optional QC Lockout feature is turned on, a control test must be completed before a patient test can be performed. The QC Lockout ensures compliance with facility policy by checking that quality control tests are run on the required (Admin User defined) frequency.

In the QC Lock Settings menu, the Admin User can define the QC Testing time intervals. The time intervals options are: No cycle, Everyday, Once a week and Once a month.



Note: For waived testing performed under a CLIA Certificate of Waiver, quality control checks must be performed by following manufacturer's instructions. Coag-Sense User Instructions recommends testing one set of Controls (High and Low) per new Lot# of Patient Test Strip Kits immediately upon receipt of shipment. All boxes with the same lot # from that shipment are then qualified and "QC Done" can be written on the top of all the boxes. Some States and Accrediting Agencies may have additional requirements or regulations for waived QC testing frequency. The meter does not require the running of controls for any calibration.

If the Control tests (High Control or Low Control) do not produce a result, or if the result is outside of the target value range, the meter is locked from further patient testing until a valid QC result is obtained. The Test button will be grayed with a lock symbol indicating that it is in QC Lockout mode.

The Admin User when logged in, can override the QC Lockout function either by disabling the QC lock feature or by changing the QC testing time interval. This will allow patient testing to proceed, and the Test button will turn blue again. Standard Users cannot change the QC Lock Setting, only the Admin User can change the settings.

8. Performing a Control Test

There are 2 Low Control Strips, 2 High Control Strips and a Control Activation Solution shipped with each test strip kit. Controls should be tested immediately upon receipt of each new lot number. Extra Controls may be ordered separately if more frequent QC testing is required.


Note: The following directions are for running a low control strip. When this procedure is complete, run a high control strip. The controls may be run in any order. The meter will display and store the results in PT seconds only. The meter does not use or require results from the control strips prior to running a patient test strip. If multiple boxes of test strips are purchased at the same time and they have the same lot number, only one low and one high control from that lot needs to be tested.

Note: If an error message appears, consult the 'Troubleshooting' section of this manual.










WARNING: DO NOT move or touch the meter while it is running a test. Unreliable results may occur.


Follow these steps to perform a test on a Low or High Control.


No.	Action	Image guided instruction
1.	Power the meter ON by pressing and holding the  (POWER) button on the right side of the meter.	

Note: The message field on the first screen will display errors encountered during self-check if any.

No.	Action	Image guided instruction
2.	Press the Login icon. User is now directed to enter login credentials or scan operator barcode using a standard barcode scanner. <i>Note: To configure your meter for Auto login please contact technical support.</i>	
3.	Successful login directs the user to the Home screen. Press the Control icon on display icon.	
4.	Select from the following two options as applicable. Low Control Test or High Control Test.	
5.	Strip lot confirmation screen displays the Lot information of the strip that was last recorded. Proceed with testing if the control strip is from the same lot. Otherwise, press Change and scan the NFC tag against the NFC Tag scanner on the meter, the Lot # (six-digit number) and Barcode # (eight-digit number) will auto populate.	Representative example only 

No.	Action	Image guided instruction
	If the NFC tag information is not available, you may manually enter the Lot # and Barcode # using the keypad on the touchscreen. Press the “Forward arrow.” Note: Ensure the strip expiration date on the strip packaging has not passed. Contact your Coag-Sense distributor for help with reordering.	 
6.	Open the packaging of the selected control strip by tearing the notched end.	
7.	Holding the round end, gently push the strip completely into the meter. The strip fits snugly when pushed all the way toward the back wall of the strip insertion area.	
8.	The meter warms the strip (for 25 seconds) to operating temperature, the display shows a countdown in seconds. Note: Do not apply the control activation solution until the warm-up is complete and the meter display shows 'Apply Control Solution'.	

No.	Action	Image guided instruction
9.	Open the control activation solution and hold at an angle to allow insertion of the transfer tube. Insert transfer tube into control activation solution. Let capillary action fill until solution flow stops at green band.	
10.	The meter beeps once and displays “Apply Control Solution” when it is ready for the Control Activation Solution. Note: You now have up to 2 ½ minutes to apply the activation solution to the control strip. Insert transfer tube tip into sample application well of test strip, touching tip down at flashing green light in front of wheel. Depress plunger completely to dispense the activation solution.	
11.	When the control activation solution is properly applied and detected, the flashing green light will turn off, and the meter displays ‘Testing Please Wait’. Note: If this screen is not displayed within 8 seconds not enough solution was applied. Remove the strip. Retest with a new control strip. DO NOT attempt to add more solution to the strip.	

No.	Action	Image guided instruction
12.	When testing is complete, the Pass/Fail results are displayed in PT units. Date and Time are also displayed. Remove and discard the Control Strips. <i>Note: Control test results only display PT seconds, this is to avoid confusing control strip INR results with patient test strip INR results.</i>	
13.	Repeat Steps 3-13 for 'High control strip.'	
14.	You can now proceed to testing patient blood samples. If you are not going to test, turn off the meter by pressing and holding the POWER button. The opened Control Activation Solution may be used until the expiration date.	

Note: If control test fails, repeat the test with a new strip. If the control test continues to FAIL, please contact your POC testing administrator or Coag-Sense Technical Support for assistance.

9. Collecting a Fingerstick Sample

Tips for a Successful Fingerstick

- Make sure that you have all the supplies needed before you start.
 - 21g Lancet device (Single use, auto disabling)
 - Sample Transfer Tubes
 - Sterile alcohol prep pads
 - Gauze square and Band-Aids
 - Biohazard waste container (SHARPS)
- For fingerstick blood testing, increasing the flow of blood in the finger will help you capture a good drop of blood. Before you perform the fingerstick, have the patient warm their hand by washing it in warm water, or by using a hand warmer. Ensure that the patient's hand is dry prior to testing.
- Do not use fingers with tight rings, scars, calluses, or other features that prevent getting good access to the blood sample.
- One of the middle or index fingers on either hand is recommended.
- Gently squeeze or massage the finger to be lanced, near the tip. Good circulation can be seen if the patient's fingertip changes to a pinkish shade.
- Use a **21g 1.8-2.4 mm** depth single-use auto-disabling lancet. **Smaller gauge/shallow depth lancets (i.e. diabetes 23g lancets) should not be used.** Refer to the Lancet device instructions for more information on use.
- Lance the fleshy part of the fingertip just slightly left or right of the center. **Press lancet firmly against finger.**
- For better blood flow, you may have the patient hold their hand below their heart. **If necessary, squeeze the finger from the sides to open the wound for proper blood flow to produce a pea sized drop.**

The best test sample is when:

- The blood is collected right after the fingerstick and put into the sample well without delay. If there is any delay in sample collection or application, repeat with a fresh fingerstick and a new strip.
- There are no bubbles or air pockets in the tube or sample.



WARNING: Patient samples, controls, used test strips, transfer tubes and lancets are potentially infectious. Dispose the strips and collection devices using universal precautions.

No.	Action	Image guided instruction
1.	Have patient wash their hands with soap and warm water. Dry completely. If using an alcohol wipe, the finger must be wiped dry with sterile gauze (air drying is insufficient to remove residual alcohol in time) <i>Note: Residual alcohol or water will affect results. Be certain that finger is completely dry. Do not use antiseptic wipes.</i>	
2.	Choose a site near the top fleshy part of the fingertip just slightly left or right of the center. Any one of the middle fingers can be used to lance. <i>Note: Avoid the more sensitive area in the center. Avoid any calluses or scars.</i>	




No.	Action	Image guided instruction
3.	Remove the cap from the single use lancet. Place it against the skin. Holding the body of the lancet, push down firmly against the finger to lance the surface of the skin. Do not lance finger until meter displays “APPLY SAMPLE.” A minimum of 10µl of collected blood sample is required. <i>Note: The blood should flow freely. If it doesn't, gently squeeze the finger to get it started. Lowering the patient's hand and arm so that the fingertip is below the heart helps the blood drop form.</i>	
 WARNING: Squeezing the fingerstick site excessively (milking) releases interstitial “tissue layer” fluid that cause unreliable results.		
4.	When ready to collect the drop of blood, hold the Sample Transfer Tube horizontal. Touch tip to bead of blood and let capillary action fill until blood flow stops at green band. Squeeze finger to generate additional blood if required to completely fill to green band.	
5.	Once you have collected the sample, IMMEDIATELY put it into the sample well on the test strip. See ‘Performing a PT Test’ section of this manual.	
 WARNING: If there is a bubble or an air pocket present in the blood sample in the transfer tube, start the test over with a fresh fingerstick on a different finger.		

10. Performing a PT Test







WARNING: Place the meter on a stationary, level surface for testing. DO NOT move the meter or allow it to vibrate during the test. Unreliable results may occur. Wear gloves and follow all applicable hygiene and safety procedures.




Follow the below steps to perform a patient test:



No.	Action	Image guided instruction
1.	Power the meter ON by pressing and holding the  (POWER) button on the right side of the meter.	
2.	Press the Login icon. Operator is now directed to enter login credentials or scan operator barcode using optional Barcode Scanner.	
3.	Successful login directs the user to the Home screen. Press the Test icon on display screen.	

Note: The message field on the first screen will display errors encountered during self-check if any.

Note: When the QC Lockout is ON, if the control test fails, then the patient test feature will be disabled, please contact your Admin for assistance.

No.	Action	Image guided instruction
4.	Patient Strip lot confirmation screen displays the Lot information of the strip that was last recorded. Proceed with testing if the test strip is from the same lot. Otherwise, press Change and scan the NFC tag against the NFC Tag scanner on the meter, the Lot # (six-digit number) and Barcode # (eight-digit number) will auto populate. If the NFC tag information is not available, you may manually enter the Lot # and Barcode # using the keypad on the touchscreen. Press the forward arrow. <i>Note: Make sure the expiration date on the strip packaging has not passed. Contact your Coag-Sense distributor for help with reordering.</i>	  
5.	If Patient ID feature is enabled, then Patient ID information must be scanned using the optional barcode scanner or manually entered in the field. Press the forward arrow	
6.	Open the packaging of the test strip by tearing the notched end.	

No.	Action	Image guided instruction
7.	Holding the round end, gently push the strip completely into the meter. The strip fits snugly when pushed all the way toward the back wall of the strip insertion area.	 05/09/2018 11:20 am Patient test Strip Lot # 111111 Please insert a strip
8.	The meter warms the strip (for 25 seconds) to operating temperature. The display shows a countdown in seconds. <i>Note: Do not apply test sample until the warm-up is complete and the meter display shows 'Apply sample.'</i>	 05/09/2018 11:21 am Patient ID aa Please wait until warm-up is complete 25
While the meter is warming up, get ready to perform a fingerstick. See "Collecting a Fingerstick Sample" section in this manual.		
9.	When the warm-up is complete, the meter beeps (if sound is turned ON) the screen displays a 'Apply Sample' message. Perform the Fingerstick. <i>Note: You now have up to 2 ½ minutes to perform a fingerstick and apply the sample to the test strip.</i>	 08/31/2018 3:13 PM Patient ID aa Apply sample
10.	IMMEDIATELY after collecting the patient sample, place the tip of the sample transfer tube at a 45° angle into the sample well on the test strip in front of the wheel where you see the flashing green light. Gently touch the tip down onto the sample well. Depress the plunger completely to dispense blood sample.	

No.	Action	Image guided instruction
	<i>Note: Discard the sample transfer tube in a biohazard container.</i>	
11.	When the sample is detected, the meter displays a 'Testing Please Wait' message. <i>Note: If this screen is not displayed, either not enough blood sample was applied, or the sample had air bubbles in it. Remove the strip and retest with a new strip and fresh fingerstick</i>	 The screenshot shows a mobile device screen with a status bar at the top displaying '05/09/2018 11:21 am' and various icons. Below the status bar is a blue header bar with 'Patient ID' on the left and 'aa' on the right. The main area of the screen is light gray and displays the text 'Testing Please Wait' in a bold, black font.
12.	When testing is complete, the meter beeps (if sound is turned ON). The results (INR and PT seconds) are displayed on the screen along with the date and time of the test. <i>Note: Memo field allows user to make notes along with the results. Upon clicking the Check mark icon, the main screen is displayed.</i> <i>Note: Refer to the "Troubleshooting" section of this manual if the meter displays messaging, like for example: CLOT TIME TOO SHORT or NO CLOT DETECTED.</i>	 The screenshot shows a mobile device screen with a status bar at the top displaying '05/09/2018 11:22 am' and various icons. Below the status bar is a blue header bar with 'Strip Lot #' on the left and '111111' on the right. The main area of the screen is light gray and displays the date and time '05/09/2018 11:21 am' at the top. Below this, there are two large white boxes with black text: '2.7' with 'INR' below it, and '33.5' with 'PT(Sec.)' below it. At the bottom of the screen, there is a white bar with a blue checkmark icon on the left, the word 'memo' in the center, and a gray circular button with a white checkmark icon on the right.
13.	Remove the test strip and dispose in a biohazard collection container. <i>Note: Repeat the test if the results seem unusually low or high. If the results still seem unusual after a second test, contact Technical Support.</i>	
14.	User may export recorded results into the electronic medical record system by connecting to a computer or middleware system.	

No.	Action	Image guided instruction
	User may also print the results as well, refer to the “Printing” section in this manual. <i>Note: The meter stores 2000 patient test results in memory with the time and date stamp. Refer to “Reviewing the Memory” in this manual for more information.</i>	
15.	Turn the meter OFF by pressing and holding the ⏻ POWER button when you are finished testing.	

**WARNING: Unexpected results**

An unexpected result may include any result that falls outside the patient's therapeutic target range, or a result that falls inside the target range but is not consistent with the patient's current health status (e.g., patient is experiencing bleeding or bruising).

What can cause unexpected results:

Certain prescription drugs (for example, heparin) and certain over-the-counter medications can affect the action of oral blood thinners and the INR value.

Changes in diet, lifestyle, or taking nutritional supplements such as ginkgo biloba can affect the action of oral blood thinners and the INR value. Liver diseases, congestive heart failure, thyroid dysfunction, Lupus, antiphospholipid antibody syndrome (APS) and other diseases or conditions can affect the action of oral blood thinners and the INR value.

Be sure to confirm whether the patient has any of these conditions before you begin testing, and any time there are changes in patient health status or medications after you have begun testing.

What to do when you get an unexpected result:



Follow instructions for re-testing on the Coag-Sense PT/INR meter. For unexpected results, contact **Technical Support at 1-866-903-0890**. Consider re-testing using an alternative method prior to adjusting the patient's dose of anticoagulant medication, or any other corrective actions.





11. Performing a STAT Test



A Stat Test allows the user to perform and obtain results of a patient test without logging in a User account or entering a Patient ID. Stat test results are saved in the memory of the meter with Patient ID as Stat, along with patient result it also records the date and time the test was performed.

Note: The memory does not store patient ID or operator ID for any Stat test performed.

Follow the steps below to perform a STAT test:

No.	Action	Image guided instruction
1.	Power the meter ON by pressing and holding the  (POWER) button on the right side of the meter. Note: The message field on the first screen will display errors encountered during self-check if any.	
2.	Press the Stat icon. User is now directed to the strip lot confirmation screen. The screen displays the Lot information of the strip that was last recorded. Proceed with testing if the test strip is from the same lot as displayed. Otherwise, scan the NFC Tag against the NFC Tag scanner on the meter, the Lot # (six-digit number) and Barcode # (eight-digit number) will auto populate If NFC Tag information is not available, you can manually enter the Lot # and Barcode # using the keypad on the touchscreen into the	 

No.	Action	Image guided instruction
	respective fields. Press the forward button. Note: Make sure the expiration date on the strip packaging has not passed. Contact your Coag-Sense distributor for help with reordering.	
3.	Open the packaging of the test strip by tearing the notched end.	
4.	Holding the round end, gently push the strip completely into the meter. The strip fits snugly when pushed all the way toward the back wall of the strip insertion area.	
5.	The meter warms the strip (for 25 seconds) to operating temperature. The display shows a countdown in seconds. Note: Do not apply test sample until the warm-up is complete and the meter display shows 'Apply Sample'.	
6.	While the meter is warming up, get ready to perform a fingerstick. See "Collecting a Fingerstick Sample" in this manual.	
7.	When the warm-up is complete, the meter beeps (if the sound is turned ON) and the screen displays a 'Apply Sample' message. Perform fingerstick now. Note: You now have up to 2 ½ minutes to perform a fingerstick and apply the sample to the test strip.	




No.	Action	Image guided instruction
8.	IMMEDIATELY after collecting the patient sample, place the tip of the sample transfer tube at a 45° angle into the sample well on the test strip in front of the wheel where you see the flashing green light. Gently touch the tip down onto the sample well. Depress plunger completely to apply blood sample. Discard the sample transfer tube in a biohazard waste container.	
9.	When the blood sample is detected, the meter displays a 'Testing Please Wait' message. <i>Note: If this screen is not displayed, either not enough blood sample was applied, or the sample had air bubbles in it. Remove the strip and retest with a new strip and fresh fingerstick.</i>	
10.	When testing is complete, the meter beeps (if the sound is turned ON). The results (INR and PT in seconds) are displayed on the screen along with date and time of the test. <i>Note: Memo field allows user to make notes along with the results. Upon clicking the Check mark icon, the main screen is displayed.</i> <i>Note: Refer to the "Troubleshooting" section of this manual if the meter displays messaging, like for example: CLOT TIME TOO SHORT or NO CLOT DETECTED.</i>	




No.	Action	Image guided instruction
11.	Remove the test strip and dispose in a biohazard collection container.	
	<i>Note: Repeat the test if the results seem unusually low or high. If the results still seem unusual after a second test, contact Technical Support.</i>	


12. Reviewing the Memory

The Coag-Sense meter stores up to 2000 patient test results and 500 control test results, along with the respective date and time of the test performed. When the memory has reached maximum storage capacity, the oldest result is automatically deleted and gets replaced with the most recent result. This meter records all test results, i.e. patient tests, control and Stat tests. To avoid the loss of stored test results, the User may export recorded results to an electronic medical record system by connecting to middleware software using POCT1-A protocol. Memory is not lost if there is a break in power for any length of time.

Follow the steps below to review results in memory:

No.	Action	Image guided instruction
1.	Power the meter ON by pressing and holding the  (POWER) button on the right side of the meter. <i>Note: The message field on the first screen will display errors encountered during self-check if any.</i>	
2.	Press the Login icon. Operator is now directed to enter login credentials or scan operator barcode using any standard Barcode Scanner.	
3.	Successful login directs the user to the Home screen. Press the Results on display screen.	

No.	Action	Image guided instruction												
4.	The 'Results' screen lists Patient test results and Control test results.													
5.	User can select from the various results stored in memory to view details of that results.	 <table border="1"> <thead> <tr> <th>ID</th><th>Date/Time</th><th>INR</th></tr> </thead> <tbody> <tr> <td>aa</td><td>05/09/2018 11:21 am</td><td>2.7</td></tr> <tr> <td>bb</td><td>05/09/2018 11:15 am</td><td>3.4</td></tr> <tr> <td>aa</td><td>04/26/2018 4:27 pm</td><td>2.9</td></tr> </tbody> </table>	ID	Date/Time	INR	aa	05/09/2018 11:21 am	2.7	bb	05/09/2018 11:15 am	3.4	aa	04/26/2018 4:27 pm	2.9
ID	Date/Time	INR												
aa	05/09/2018 11:21 am	2.7												
bb	05/09/2018 11:15 am	3.4												
aa	04/26/2018 4:27 pm	2.9												
6.	User can filter and view test results for a specific Patient Identifier by clicking on the Show Results of History button. <i>Note: User may choose to list result history for a patient by selecting Patient ID. This example displays results for Patient ID 'aa'.</i>													

No.	Action	Image guided instruction
7.	Clicking the chart icon displays the test result for the selected patient in a chart/list output.	

13. Printing

With the portable printer (optional accessory) available from Coag-Sense, results from the Coag-Sense meter memory can be printed on either thermal paper or other media such as thermal labels for applying to patient charts.


Note: Contact your Coag-Sense distributor for help with ordering.


Note: Refer to the User Manual provided with the printer for its general operation.

What you'll need:

- Coag-Sense meter
- Optional Portable Printer, Catalog # 03P76-01
- 2" Thermal Paper, Catalog # PD99906-OM

Follow the steps below for printing results.

No.	Action	Image guided instruction
1.	Plug the USB cable from the portable printer into the USB port of the meter.	
2.	From the individual result display click the printer icon on the lower-right hand corner of the screen to print the test results.	

No.	Action	Image guided instruction
3.	User may choose from three different printer connection modes. Network and Bluetooth connection modes will require configuration using the setup icon prior to printing. Press Print to print the result.	

If the results fail to print, confirm that the printer is ON and charged as it automatically turns OFF after a few minutes. Note that the meter can only print to specific printers qualified and supplied by CoaguSense. The meter cannot print directly to your computer printer. The printer can be connected via Bluetooth connectivity. Refer to Coag-Sense Prothrombin Time (PT)/INR Monitoring System Portable Printer Guide, CSI P/N 200464.

For assistance contact CoaguSense Technical Support @ support@coagusense.com or call Technical Support at 1-866-903-0890.

14. Network Connectivity and Security

The Coag-Sense meter offers the ability to send patient test results to electronic medical records or laboratory information systems. The Coag-Sense meter can be connected to a Local Area Network (LAN) via either a direct Ethernet connection or wirelessly using Wi-Fi. We offer WPA2-Enterprise level of security to log into hidden WIFI if required. If this meter is connected to a Local Area Network, the network must be protected against unauthorized access. It must not be linked directly to any other network or the Internet. Customers are responsible for the security of their Local Area Network, especially in protecting it against malicious software and attacks. This protection might include measures, such as a firewall to separate the device from uncontrolled networks and intrusion detection that ensure that the connected network is free of malicious code including intrusion detection.

If you use commercial middleware or a customized data management system solution, ensure that sensitive patient identifiable data transmitted via the POCT1-A interface is protected by appropriate security measures.

Ensure that the instrument is protected against unauthorized physical access and theft and that the tamper event label has was intact on delivery.


Do not use shared user or operator accounts on meter and network.

Whether working in a wired or wireless environment, use a strong alpha and numeric password for user or operator accounts on the meter and network. Observe your own facility guidelines on password management where available.

Report any malfunction or security risks such as unauthorized access, modification, interference or vulnerabilities that occur related to Coag-Sense PT/INR Monitoring System to Technical Support @ support@coagusense.com or call Technical Support at 1-866-903-0890.

15. Meter Software Update

When connected to the Internet, the Coag-Sense meter can check for the availability of meter software updates which can be downloaded. **If the meter is not continuously connected to the Internet, you should make it a practice to periodically connect the meter to the Internet to check for software updates.** If a critical update is available, the meter may require the installation of an update prior to proceeding with testing. Make sure to check if the battery is charged enough before performing an update. If the battery's charge is not enough and then the meter is abruptly turned off during update, an error may occur on the meter.



16. Battery

The Coag- Sense meter has a factory installed Rechargeable Lithium Polymer Battery, that charges when the power adapter (provided with the meter) is plugged into the wall socket. A charged battery (6 hours to fully charge) can run approximately 200 tests. Please be aware that battery life may be affected by many factors such as operating conditions (e.g. ambient temperature), frequency of use, and test duration.

Note: *The Battery is not User replaceable.*



When the battery is running low the status bar on the touchscreen of the meter displays a red indicator in the 'Battery status' icon. The touchscreen displays a 'Low Battery' warning. The battery begins charging as soon as the power adapter is connected to the power supply.



WARNING: Lithium Polymer batteries may explode or combust if mishandled. Do not subject the meter to prolonged exposure to sunlight or place the meter on or in heating appliances such as microwave, conventional oven or radiator. Only charge battery using the power adapter provided along with the Coag-Sense meter. Use of other power cables may result in damage to the meter. Do not disassemble or dispose battery in fire. Do not charge/discharge battery out of recommended temperature range.

17. Cleaning and Disinfecting the Meter

No maintenance is required other than routine cleaning and/or disinfecting.

When the power is off and the USB cable is not connected, the meter housing can be cleaned and disinfected. Wipe all exposed surfaces with Healthcare Bleach Germicidal Wipes containing Sodium Hypochlorite (EPA No. 67619-12) for a contact time of 1 minute to pre-clean blood and other body fluids. Caution should be taken to not get fluids inside the meter through the test strip port, data transmission port or battery compartment. Dispose of the used towelette. The meter should be allowed to air dry before use.

If instructions for use are properly followed, patients should not come in direct contact with the Coag-Sense meter thereby reducing the possible transmission of bloodborne pathogens between patients. Sample should always be transferred from the patient to the meter using a new disposable sample transfer tube.



The test strip is designed to contain the patient sample, preventing it from entering the meter. Do not clean/disinfect inside the meter where the test strip is inserted. Cleaning this area should be avoided. Please call Technical Support at 1-866-903-0890 if this area requires cleaning/disinfecting.



WARNING: Do not put the meter in liquid. Do not allow liquids to get inside the meter or into any of the connectors or plugs on the meter. If you suspect any physical damage or deterioration of the meter (such as cracking or gross distortion), or if the meter does not turn on after cleaning, call Technical Support.

Always refer to local, state and federal disinfecting guidelines. More information on bloodborne pathogen safety and proper disinfecting techniques can be found at:

"FDA Public Health Notification: Use of Fingerstick Devices on More than One Person Poses Risk for Transmitting Bloodborne Pathogens: Initial Communication" (2010)

<http://wayback.archive-it.org/7993/20170111013014/http://www.fda.gov/MedicalDevices/Safety/AlertsandNotices/ucm224025.htm>

"CDC Clinical Reminder: Use of Fingerstick Devices on More than One Person Poses Risk for Transmitting Bloodborne Pathogens" (2010)

<http://www.cdc.gov/injectionsafety/Fingerstick-DevicesBGM.htm>

18. Troubleshooting

The Coag-Sense Meter continually checks its systems for unexpected conditions. These may arise because of defective components or consumables, environmental factors or due to User handling and procedure errors. This section details how to resolve most problems that you might encounter. If you have any questions or problems during the troubleshooting process, note the display wording and contact Technical Support at **866-903-0890** or email **techsupport@coagusense.com**.

Meter Display	Possible Cause(s)	Solution
ROOM TEMP INCORRECT SEE MANUAL	The temperature of the room is either below or above the operating temperature range of the meter. Battery was depleted and then plugged into AC adapter while repeated testing continued. The charging of the battery can generate enough heat to raise the internal temperature of the meter outside the operating range.	Move the meter to a place that is within the operating temperature range of the meter (65°F to 90°F, 18°C to 32°C) and allow meter time to adjust to correct temperature. Repeat testing. Remove AC power and allow meter to cool prior to continuing testing or suspend testing until battery has charged and the internal temperature has cooled down sufficiently.
REMOVE STRIP	Meter turned off with used strip in it. If no strip present, possible shipment damage.	Remove the strip and begin again. Call Technical Support.
WHEEL PROBLEM	The test strip was not inserted fully or may have been inserted at an incorrect angle or incorrect speed. There may be a problem with the wheel on the strip or with the meter.	Reinsert the strip holding the back of the meter steady with one hand while inserting the strip completely using a quick smooth motion with the other hand. If display persists, try again with another new strip. If the message displays again contact Technical Support.

Meter Display	Possible Cause(s)	Solution
DETECT PROBLEM	There may be a problem with the strip insertion or with the motor carriage in the meter.	Take the strip out and reinsert holding the back of the instrument steady with one hand while inserting the strip completely with the other hand. Insert the strip using a quick smooth motion. Try again with another strip. If the message persists, contact Technical Support.
LIQUID PROBLEM	There may be a problem with the strip or with the optical system of the meter.	Take the strip out and reinsert holding the back of the instrument steady with one hand while inserting the strip completely with the other hand. Insert the strip using a quick smooth motion. Try again with another strip. If the message persists, contact Technical Support.
MOTOR PROBLEM	There may be a problem with the motor function of the meter.	Turn the meter off then back on. Try again with another strip. If the message persists, contact Technical Support.
HEATER PROBLEM	The meter is too warm, too cold, or there may be a problem with the meter.	Move the meter to a place that is within the operating temperature range of the meter (65°F to 90°F, 18°C to 32°C) and allow meter time to adjust to correct temperature. Repeat testing. Turn meter off then on again after 5-7 minutes Try again with another strip. If the display persists, contact Technical Support.

Meter Display	Possible Cause(s)	Solution
TEST STRIP EXPIRED SEE MANUAL	The lot of strips have expired. Meter date is not set correctly.	Use a different lot of strips that has not expired. Verify the date setting on the meter is current.
NO SAMPLE DETECTED	Either no sample or not enough sample was applied to the strip within 2 1/2 minutes after the "Apply Sample" message was displayed. This can also happen if sample is applied on the strip but outside of the sample application well.	Repeat the entire procedure (including fingerstick on a different finger) with a new strip. Apply the sample within 2 1/2 minutes after display of the "Apply Sample" message. Ensure that the transfer tube is filled to the white plug and touches the sample well before dispensing sample.
CONTROL FAIL-NO CLOT DETECTED	There was no clot formation; sample clotting time was very long and out of testing range. There was insufficient control activation solution transferred to the test strip. Possible causes include; an air bubble in the sample or not allowing control activation solution to completely fill transfer tube. This may be due to a problem with the shipment/storage of the control strips or the control activation solution. Plasma on control strips is sensitive to exposure to temperatures outside the storage range.	Repeat the entire procedure with a new control strip. If the same message persists and if you have additional inventory of the test strip kit from the same kit lot, use the control strip from that box(es). If you don't have additional inventory OR if the error message persists, contact Technical Support. Control strips should be tested immediately upon receipt of your shipment of new test strips as they have a limited shelf life. This does not indicate meter malfunction
CONTROL FAIL-CLOT TIME TOO SHORT	The clotting time was very short and out of testing range (<8 seconds).	Repeat the entire procedure with a new control strip. Visually confirm that no air bubbles are in the control

Meter Display	Possible Cause(s)	Solution
	An air bubble was detected in the control activation solution sample. The sample transfer tube was not filled with the control activation solution to the white plug Applying the control activation solution to the test strip before "Apply Control Solution" displayed on screen.	activation solution sample before applying to test strip. Ensure that the tube is filled to the white plug. Depress black plunger completely to dispense the control activation solution sample. Repeat the entire procedure with a new strip. If the same message persists and if you have additional inventory of the test strip kit from the same kit lot, use the control strip from that box(es). If you don't have additional inventory OR if the error message persists, contact Technical Support.
CONTROL FAIL-OUT OF RANGE	The control strip result is outside of its acceptable range (FAIL- out of range). This may be due to a problem with the shipment/storage of the control strips or the control activation solution. Plasma on control strips has a limited shelf life and the clotting time will change when exposed to temperatures outside the storage range.	Repeat test with another control strip. If the second test is out of range, contact Technical Support. Control strips should be tested immediately upon receipt of your shipment of new test strips as they have a limited shelf life. This does not indicate a meter malfunction.

Meter Display	Possible Cause(s)	Solution
CLOT TIME TOO SHORT	The clotting time was very short and out of testing range (<8 seconds). An air bubble was detected in the sample. Lancing the finger before "Apply Sample" displayed on screen. Taking too long to collect the sample in transfer tube (make sure of using 21g needle lancet for good flow of blood flow).	Repeat the entire procedure (including fingerstick on a different finger) with a new strip. Visually confirm that no air bubbles are in the sample before applying to test strip. Depress black plunger completely to dispense the sample. If the same message repeats, contact Technical Support.
NO CLOT DETECTED	The sample clotting time was very long and out of testing range. There was insufficient sample transferred to the test strip. Possible causes include; improper lancing (21g lancet required), an air bubble in the sample or not allowing sample to completely fill transfer tube.	Confirm that the patient has not recently taken heparin or other contraindicated drugs listed on the test strip package insert. Visually confirm that no air bubbles are in the sample before applying to test strip. Depress black plunger completely to dispense the sample Repeat the entire procedure (including fingerstick) with a new strip. If the same message displays, use an alternative testing method and contact Technical Support.
LOW BATTERY	The meter battery is low	The meter can complete the current test. The meter connected to the power

Meter Display	Possible Cause(s)	Solution
		adapter should be plugged into the wall socket.

General Troubleshooting

Issue	Possible Causes	Solution
Meter does not power ON	Insufficient Battery to Power ON. Power Adapter not connected properly for charging the battery. Not pressing and holding Power button when turning meter on.	Check if the power adapter (provided with the meter) is connected to the port in the Meter and the wall socket. If issue persists, or if the power adapter is faulty, contact Technical Support.
Cannot insert strip completely	Accumulation of dirt, dust, control activation solution, or blood in the strip insertion area. Wheel is not seated properly into of test strip.	Contact Technical Support for assistance with cleaning the strip insertion area. Use your thumbnails to push wheel spindles down to snap wheel into place. If issue persists, Contact Technical Support.
Touch screen display issues	Insufficient/Low Battery Display Faint or low brightness	Connect the power adapter to wall socket. Change 'Brightness' User setting in the 'Device Settings' menu of the meter.

Issue	Possible Causes	Solution
Touch screen not responding	Prolonged exposure to direct sunlight.	Avoid prolonged exposure to direct sunlight, as it may reduce life expectancy and functionality of the display.
	Dropping or subjecting the meter to strong shocks.	Contact Technical Support.
Touch screen scratched or cracked	Dropping or subjecting the meter to strong shocks	Contact Technical Support.
	Using pointed or sharp-edged objects other than the recommended 'finger' or rubber stylus to touch the screen elements.	
Power Adapter Not working	Faulty Adapter (Bent power cord, Bent power pin in the meter)	Check adapter functionality by plugging the power adapter to a different wall socket.
		If issue persists, contact Technical Support.
Software Issues	Software version update issue	Power cycle and re-install new software version if available (<i>Settings_Device Information settings_Software Version</i>).
		If issue persists, press Reset button to restore factory settings. If issue still exists, contact Technical Support.
Lost NFC Tag	Misplaced NFC Tag	The NFC tag is affixed to each test strip kit box. Otherwise, enter the strip information manually into the touchscreen to perform the current test.
		Alternately, if you have additional inventory of the test strip kit from the same kit lot, use the NFC tag from that box(es).

Issue	Possible Causes	Solution
NFC Tag Issues: • NFC tag not working • Scanned information does not match the information on the strip packaging	Improper scanning of the NFC tag. Faulty NFC Tag scanner in the meter	Touch or bring the NFC tag to proximity of the NFC Tag scanner. If the issue persists enter the strip information manually into the touchscreen to perform the current test. If you have more than one meter, try scanning the NFC Tag on another meter to narrow down the root cause to either the tag or scanner. If the issue persists, contact Technical Support.
NFC Tag scanner issue	Tag Scanner works intermittently or does not work. Scanned NFC scan did not match the Lot # and Barcode # on the test strip.	Scan the alternate NFC tag provided. Touch or bring the NFC tag to proximity of the NFC Tag scanner. If the issue persists enter the strip information manually into the touchscreen to perform the current test. If you have more than one meter, try scanning an NFC Tag on another meter to narrow down the root cause to either the tag or scanner. If the issue persists, contact Technical Support.

19. Warranty

Limited One (1) Year Warranty

Use of the Coag-Sense PT/INR System

The Coag-Sense PT/INR system is designed for use in monitoring patients on oral anticoagulant therapy. Proper adherence to the instructions in this User Manual and package insert are critical to proper operation.

WARNING: Failure to comply with the User Manual could lead to inaccurate PT/INR results which could lead to incorrect medication dosing which could lead to injury or death.

Limited Warranty

CoaguSense warrants that the Coag-Sense meter is free from all defects in material and workmanship for a period of one (1) year from date of purchase. When the meter is used for the intended purpose and in the appropriate manner, the remedy is repair or replacement at CoaguSense's option. The warranty does not apply to a meter damaged by misuse, alteration or tampering to either the hardware or software. Contact Technical Support at 1-866-903-0890 for instructions.

THIS WARRANTY APPLIES ONLY TO THE METER. COAGUSENSE'S ENTIRE LIABILITY IN CONNECTION WITH THE METER, REGARDLESS OF THE LEGAL OR EQUITABLE BASIS OF ANY CLAIM, IS LIMITED TO THE PURCHASE PRICE OF THE METER. IN NO EVENT SHALL COAGUSENSE, INC. BE LIABLE TO THE PURCHASER FOR ANY INCIDENTAL, CONSEQUENTIAL (INCLUDING BUT NOT LIMITED TO LOSS OF INCOME OR PROFITS) SPECIAL, INDIRECT, OR PUNITIVE DAMAGES ARISING FROM OR IN ANY WAY CONNECTED WITH THE PURCHASE OR OPERATION OF THE METER OR ITS PARTS. NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE IS IMPLIED FROM THE SALE OF THE COAG-SENSE PT/INR SYSTEM. NO WARRANTY, EXPRESS OR IMPLIED (IF ANY) SHALL EXTEND FOR A LONGER DURATION THAN THE DURATION OF THE EXPRESS WARRANTY STATED ABOVE.

Instructions for Meter or Product Return

Upon review and agreement with CoaguSense Technical Service, you may be directed to return the unit. Should this occur, clean the outside surface as described in the “Cleaning and Disinfecting the Meter” section. The original packaging may be required for this purpose.

20. Reordering Information

For a description of the products listed below, please see the information above.

Product	Catalog #
Coag-Sense PT/INR Professional System	03P70-02
Coag-Sense Test Strip Kit, Box of 50	03P56-50
Coag-Sense Control Strip Kit -10	03P69-10
Sample Transfer Tubes with Preloaded Plungers, 1 pack	03P52-55
Single-Use, 21g 2.2mm depth Auto Safety Lancets - Box of 100	03P58-04
AC Power Adapter – U.S.	03P74-01
Optional Portable Printer	03P76-01
Thermal Printer Paper, 2"	PD-99906
Replacement Carrying Case	03P75-01
Stylus Pen	03P77-01

21. EMC Tables

The following tables contain the Manufacturer's declaration and additional information required by IEC60601-2:2014 (Fourth Edition).

Test Name	Ref. Standard	Ports to Test	AC Mains Voltage	Test Level Required	Notes
Mains Terminal Disturbance Voltage	CISPR 11:2015± A1:2016	AC Mains	100V-50Hz 100V-60Hz 220V-60Hz 230V-50Hz	Group I, Class A	
Radiated Disturbance	CISPR 11:2015± A1:2016	Enclosure	100V-50Hz 100V-60Hz 230V-50Hz 220V-60Hz 230V-50Hz	Group I, Class A	
Harmonic Current Emissions	EC 61000-3-2:2014	AC Mains	230V-50Hz	Class A	
Voltage Fluctuations & Flicker	EC 61000-3-3:2013	AC Mains	230V-50Hz	Pst = 1 Pit = 0.65 Dmax = 4 DC = 3.3%	
Electrostatic Discharges (ESD)	EC 61000-4-2:2008	Enclosure	230V-50Hz	±8kV Contact ±2kV, ±4kV, ±8kV, ±15kV Air	1 pulse/ 1sec contact 8kV air 15kV
Radiated RF Electro-Magnetic	EC 61000-4-3:2008± A1-2007±		230V-50Hz	10V/m	Dwell time is 3 sec

Test Name	Ref. Standard	Ports to Test	AC Mains Voltage	Test Level Required	Notes
Fields	A2.2010	Enclosure		80MHz to 2.7GHz 80%AM at 2Hz RF Wireless Comm. (Refer to test report clause 1..15)	
Electric Fast Transients & Burst	EC 61000-4-2:2012	AC Mains	230V-50Hz	±2kV AC, 100kHz PRR	
		I/O Lines>3m		±1kV AC, 100kHz PRR	
Surges	EC 61000-4-5:2014/AMD:2017	AC Mains	230V-50Hz	±0.5kV, ±1kV L1 to L2 (DM)	5 pulses at 0° 90° 180° 270°
Conducted Disturbances, Induced by RF fields	EC 61000-6:2013	AC Mains & all I/O	230V-50Hz	3Vrms 150kHz to 80MHz 6Vrms in ISM and Amateur radio bands between 0.15MHz & 80MHz	Dwell time is 3 sec
Voltage Dips, Interruptions & Variations	EC 61000-411:2004/AMD:2017	AC Mains	100V-60Hz 240V-60Hz	0% UT for 0.5 cycle	At 0°, 45° 90° 135° 180° 225°

Test Name	Ref. Standard	Ports to Test	AC Mains Voltage	Test Level Required	Notes
					270° 315°
				0% UT for 1 cycle 60Hz: 70% UT for 30 cycles	At 0°, 180°
				60Hz: 0% UT for 300 cycles	At 0°, 180°
Power Frequency Magnetic Field	EC 61000-8:2009	Enclosure	230V-50Hz	30A/m	

22. Index

A

Admin Settings, 26
Alarm, 23
Alcohol Wipe, 7, 35

B

Battery, 5, 62
Battery Low, 54
Blood Flow, 34
Bluetooth Setting, 25
Brightness, 23

C

Calibration, 28
Chlorhexidine gluconate, 14
Class II Equipment, 17
Cleaning the Meter, 55
Clot Time Too Short, 40
Collecting a Fingerstick
Sample, 34
Communication Setting, 24
Control Activation Solution, 5
Control Fail No Clot Detected,
59
Control Strips, 5
Customer Service, 2, 73

D

Date, 8
Date and Time Setup, 22
Device Settings, 21
Direct oral anticoagulants
(DOACS), 13

E

E-Mail Support, 73
EMC Tables, 68
Ethernet Port, 4
Expected Values, 12
External Server Connection,
26

F

Fingerstick Sample, 34

H

Hazards And Symbols, 17
Heater Problem, 58
High Control Strip, 4, 29
Home Screen, 4, 9

I

IEC 60601-1, 15
INR, 5
INR Results, 38, 43
International Normalized Ratio
(INR), 12
Isopropyl Alcohol (IPA), 7, 13

L

Lancet, 7, 34, 61
Language Setup, 22
Liquid Problem, 58
Low Control Strip, 7, 29

M

Measuring Range, 12
Memory, 10, 48
Meter does not power ON, 62
Meter Setup, 19

Meter Specifications, 10
Motor Problem, 56

N

Near Field Communication
(NFC) Tag, 4
Network Connectivity and
Security, 52
No Sample Detected, 59
Normal Range, 12

O

Operating Conditions, 19

P

Performance Characteristics,
12
Performing a Control Test, 29
Performing a Patient Test, 35
Power ON/OFF, 4
Power Supply, 6
Printing, 50
PT Seconds, 4, 29, 33, 40

Q

QC Lock Settings, 26
QC Lockout, 27

R

Reset Button, 4
Room Temperature, 19

S

Sharps Container, 7, 34
Sleep Settings, 26
Software Update, 53
Software Version Update, 21
Sound Setting, 23
Squeezing the Fingertick
Site, 13

STAT Test, 43
System and User Settings, 20
System Description, 3

T

Time Zone Setup, 25
Troubleshooting, 57

U

Unexpected Results, 42
User Accounts, 27
User Management, 27

W

Warnings and Precautions, 13
Warranty, 65
Wheel Problem, 57
Wired Setting, 25
Wireless (Wi-Fi) Setting, 24

Technical Support

CoaguSense, Inc.

48377 Fremont Blvd., STE. 113

Fremont, CA 94538 USA

Toll Free: 1-866-903-0890

E-Mail: techsupport@coagusense.com



CoaguSense, Inc.
48377 Fremont Blvd., STE. 113
Fremont, CA 94538 USA
Tel: 1-510-270-5442
Fax: 1-510-226-6540
www.coag-sense.com



EMERGO EUROPE
Prinsessegracht 20
2514 AP, The Hague
The Netherlands



CSI P/N 200220 Rev AE