

Hold-n-Place® Foley Catheter Tube Holder

Leg Band Application


1. Position leg band under and high around patient's thigh with Dale product label pointing to inside of leg. Stretch leg band equally from both sides into place and fasten hook tab. Proper compression has been achieved when two fingers fit snugly under the band. 2. Place Foley catheter over white



stitching of green locking device. Leave an ample loop in catheter above leg band to avoid bladder traction: or stretch catheter for traction when required. 3. With catheter in desired position, insert the narrow green tab over



opening so the green tab overlaps. 4. Lock catheter in position by pulling green tabs in opposite directions Secure tabs in place. To reposition catheter: raise either green tab, adjust catheter, and refasten tab.



Another method of securement is to locate the Y port of the proximal end of the catheter into the locking



Waist Band Application

1. To secure the catheter in place position the waist band around the abdomen, above the wound site. No shaving necessary. The green catheter locking tab should be centered over the drain site. After the band has been positioned, press locking tab onto the waist band.



- 2. Place catheter over white stitching of green tab.
- 3. Repeat steps 3,4, & 5 above.

Additional Benefits

To help prevent meatal irritation, leg band can be rotated in either direction. Alternating legs periodically may also add to greater patient comfort.

For single patient use. Disposable

For an application video please visit www.dalemed.com

Dale medical products are available from your preferred hospital/medical supply dealer.

Not made with natural rubber latex.

Dale & Hold-n-Place are registered trademarks of Dale Medical Products, Inc. in the USA, EU, and UK





Dale Medical Products, Inc. 40 Kenwood Circle • Suite 7

Franklin, MA 02038-3298, USA T: +1-508-695-9316 www.dalemed.com





Schiffgraben 41 D-30175 Hannover, Germany





MDSS CH GmbH Laurenzenvorstadt 61 5000 Aarau, Switzerland